

Temperatur		Widerstand			max. Fehler	Temp. Koeff.
°C	°F	R_{typ}	R_{min}	R_{max}	°C	%/K
-55	-67	980	951	1009	±3.02	0.99
-50	-58	1030	1000	1059	±2.92	0.98
-40	-40	1135	1105	1165	±2.74	0.96
-30	-22	1247	1218	1277	±2.55	0.93
-20	-4	1367	1338	1396	±2.35	0.91
-10	14	1495	1467	1523	±2.14	0.88
0	32	1630	1603	1656	±1.91	0.85
10	50	1772	1748	1797	±1.67	0.83
20	68	1922	1901	1944	±1.41	0.80
25	77	2000	1980	2020	±1.27	0.79
30	86	2080	2057	2102	±1.39	0.78
40	104	2245	2217	2272	±1.64	0.75
50	122	2417	2383	2451	±1.91	0.73
60	140	2597	2557	2637	±2.19	0.71
70	158	2785	2737	2832	±2.49	0.69
80	176	2980	2924	3035	±2.80	0.67
90	194	3182	3118	3246	±3.12	0.65
100	212	3392	3318	3466	±3.46	0.63
110	230	3607	3523	3691	±3.93	0.59
120	248	3817	3722	3912	±4.70	0.53
125	257	3915	3815	4016	±5.26	0.49
130	266	4008	3901	4114	±6.00	0.44
140	284	4166	4049	4283	±8.45	0.33
150	302	4280	4153	4407	±14.63	0.20